



LACS



Low Asbestos Content Scheme Individual Results : Round 013 LACS13

For Laboratory Number : 1848 CLG Chemisches Labor Dr. Graser KG

From : 14/06/2021 To : 26/10/2021

Report Issued 05/11/2021

Report No. 15/393/1596

Sample	Your Result	Assigned Result	Score
1	Crocidolite Total Asbestos = 0.024%	Crocidolite Total Asbestos = 0.07%	0 Z : -2.3

HSE
Melanie Clunas
LACS Scheme Co-Ordinator
Fibres Team : Chemical & Biological Risks Unit

HSE, Science & Research Centre
Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK

FTPT F0931

ISSUE 2

Page 1 of 1

