



Low Asbestos Content Scheme Individual Results : Round 014 LACS014

For Laboratory Number : 1848 CLG Chemisches Labor Dr. Graser KG

From : 25/10/2021 To : 11/02/2022

Report Issued 28/02/2022

Report No. 17/393/1653

Sample	Your Result	Assigned Result	Score
1	Amosite Total Asbestos = 0.03%	Amosite Total Asbestos = 0.05%	0 Z : -1.18

HSE
Melanie Clunas
LACS Scheme Co-Ordinator
Fibres Team : Chemical & Biological Risks Unit

HSE, Science & Research Centre
Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK

FTPT F0931

ISSUE 2

Page 1 of 1

